

A Spectral-Scanning Magnetic Resonance Imaging (MRI) Integrated System

Arjang Hassibi^{1,2}, Aydin Babakhani¹, and Ali Hajimiri¹

¹California Institute of Technology, Pasadena, CA 91125, USA

²University of Texas at Austin, Austin, TX 78712, USA

Abstract- An integrated spectral-scanning magnetic resonance imaging (MRI) technique is implemented in a 0.12 μm SiGe BiCMOS process. This system is designed for small-scale MRI applications with non-uniform and low magnetic fields. The system is capable of generating customized magnetic resonance (MR) excitation signals, and also recovering the MR response using a coherent direct conversion receiver. The operation frequency is tunable from DC to 37MHz for wide-band MRI and up to 250MHz for narrow-band MR spectroscopy.

I. INTRODUCTION

Scaling down magnetic resonance imaging (MRI) [1] and nuclear magnetic resonance (NMR) spectroscopy systems [2] is an exceptionally challenging task. Nonetheless, its realization has always been a goal in medical tomography imaging (3D imaging) due to its huge potential applications in non-invasive imaging for the field of point-of-care (PoC) diagnostics.

To design a truly desktop and eventually hand-held magnetic resonance (MR) system, we are required to scale down all the components of the spectroscopy system, while keeping the sensitivity, detection time, spectral resolution, and the spatial resolution (relative to the sample size) equivalent to the specification of large MR systems.

The reciprocity theorem of electromagnetic theory, states that the electromotive force (emf) voltage induced in the coils of the MR system, $E(t)$, is describe by

$$E(t) = - \int_V \frac{\partial}{\partial t} [\vec{B}_n(\vec{r}) \cdot \vec{M}(\vec{r}, t)] dV, \quad (1)$$

where $\vec{B}_n(\vec{r})$ is the magnetic field generated by unit current in the coils at point \vec{r} , and $\vec{M}(\vec{r}, t)$ is the magnetic moment (and dipole) density created by the magnetization field \vec{B}_0 in the sample volume V (see Fig.1).

In scaling down by factor α , $\vec{M}(\vec{r}, t)$ and $E(t)$ are reduced by $1/\alpha^3$, since the number of signal-generating magnetic dipoles (e.g., ^1H in proton MR spectroscopy and imaging) is inversely proportional to volume. However, Ampere's Law, indicates that $\vec{B}_n(\vec{r})$ is proportional to α to the first order, indicating that smaller coils are more "sensitive". Therefore, by utilizing small-scale coils intimately close to the object, we still experience significant




Fig 1: Magnetic resonance spectroscopy system.

signal reduction, with a theoretical lower boundary of $1/\alpha^2$. This concept has been studied extensively in small volume NMR [3] and high-resolution MRI systems [4,5].

On the other hand, α has little effect on \vec{B}_0 since magnetization can be provided by sample- and coil-independent sources. For spectroscopy applications, it is common to find huge magnetization setups, typically of the super-conducting type, which are able to create up to 10-20T to analyze the MR spectrum of small objects. Such approaches are clearly impractical and costly for any PoC application. For portable MR system, it is more appropriate to use small-scale permanent rare-earth magnets which can produce up to 1.5T peak magnetic field. Their successful implementation has been previously reported for NMR spectroscopy and surface MR tomography applications [6,7].

Although permanent magnets can be assembled to have a small dimension suitable for PoC systems, the magnetic field surrounding them is very inhomogeneous and non-uniform. Hence, if we place an object containing magnetic dipoles (e.g., ^1H with gyromagnetic ratio of $\gamma=42.58$ MHz/T) in their vicinity, the dipoles within the object will demonstrate a wide range of MR frequencies proportional to their local field strength. Based on Torrey-Bloch formulations magnetic gradients significantly reduce the lifetime of the dipoles if, dipoles are subject to diffusion spreading [8]. This means that non-uniform fields generated by permanent magnets not only widen the bandwidth of MR spectrum, but also increases the decay of their MR response. Both of these impediments along with $1/\alpha^2$ to $1/\alpha^3$ degradation in the signal amplitude are fundamental barriers in scaling MR-based systems today.

In this paper, we show that by leveraging the advantages of integrated circuits, we can address some of the aforementioned challenges. This is done by designing a wide-band MRI system which uses the spectral-scanning MRI technique which exploits the non-uniformity of the magnetic field to acquire tomographic information [9]. We have designed, fabricated, and tested this system in a $0.12\mu\text{m}$ SiGe BiCMOS process.

II. IMAGING METHOD SPECIFICATIONS

The method of detection, spectral-scanning MRI implemented in our system, is based on exploiting the non-uniformity of the magnetic field to create coordinate-dependant MR frequencies within the sample [9]. For a given permanent magnet, the magnetic field is non-uniform, but still deterministic. Accordingly, we have a mapping between the coordinates of magnetic dipoles and their MR frequency, as shown in Fig. 2. This method, in principle, is a wideband spectroscopy technique which performs parallel MR analysis to get tomographic information about the magnetic dipole density within the imaging volume. This is also much in contrast with conventional MRI which is essentially a narrowband spectroscopy technique [1].

As mentioned before, to generate the magnetization field B_0 in small-scale MR systems, rare-earth permanent magnets are used. In such arrangements, the strength of the field is typically less than 1T. The field gradient $G = \nabla B_0$, can also vary from 0.01 to $40\text{T}\cdot\text{m}^{-1}$ in spatial coordinates where magnetic field strength has 10% to %80 of its maximum value at the magnet surface. To give a practical example, such a magnetization setup creates proton (^1H) MR frequency range of 4.2MHz to 34MHz and maximum transversal decay of 0.1msec within the imaging volume based on Torrey-Bloch formulations [8]. Therefore, we require a MR detection system capable of operating within approximately one frequency decade (almost 4-40MHz) and with a spot bandwidth in excess of 10kHz.

One important concern regarding operating at the 4-40MHz frequency range is the wide-band nature of the detection, which is greater than an octave and closer to a decade. This potentially creates practical challenges in both the signal generation (transmitter) and signal detection circuitry (receiver) of the spectral scanning system due to in-band high-order harmonics. Consequently, we have decided to keep the operation bandwidth of the MR system (both transmitter and receiver) within one octave in our design, however make the system architecture tunable and capable of sweeping its operational frequency to cover the whole 4-40MHz decade.

The basic idea is to simultaneously excite multiple frequencies within the range of interest and detect them with a narrowband system. This parallel analysis is very conducive to the parallel implementation of the system, and significantly increases the imaging and processing speed.




Fig 2: The block diagram of the spectral scanning MRI system, where equi-magnetic field surfaces translate to specific MR resonance frequencies.




Fig.3: Architecture of the spectral scanning MRI chip.




Fig. 4: Chip micrograph

Different equi-magnetic field shells across the sample are mapped to different MR frequencies. To resolve the spatial ambiguity within a shell, (see Fig. 2) multiple measurements are performed while changing the relative position of the object with respect to the magnet. The spectral scanning approach expedites this process substantially.

III. CIRCUIT IMPLEMENTATION AND RESULTS

We have illustrated the spectral-scanning MRI system architecture and its die micrograph in Fig. 3 and Fig. 4 respectively. The system consists of the *transmitter* which generates the customized MR excitation signal (pulses), and the *receiver* which recovers and coherently detects the MR responses, both integrated on the same die.

Based on the specifications of our system, we are required to generate an excitation spectrum with bandwidth of just short of one octave, yet tunable and capable of covering the whole 4-40MHz. In our design, we have implemented six tunable channels in the transmitter which collectively and when combined create the desired spectrum. In the transmitter (see Fig.3) the excitation tone signals for all channels are generated by using a reference frequency and digital counters working as frequency dividers. The divisors for the digital counters (8 through 13 in our design) are selected such that the output frequencies of all 6 channels stay within one frequency octave and avoid complications related to the second-order harmonics. After the tone generation using the digital dividers, we create the I and Q signals which are used in the transmitter and in the direct conversion receiver. The I and Q generator essentially is a divide by 4 counter circuitry, which results in an overall frequency division of {32,36,40,44,48,52} in the channels. In Fig. 5 we have illustrated the frequency allocation of each channel within the output spectrum.

Generally in MRI and NMR systems, the excitation sinusoidal pulses have a short duration with time-varying amplitude specifically designed based on the imaging technique and image reconstruction algorithm [1]. Accordingly, to satisfy this requirement in the signal path of each channel we implement a synchronized enabling signal which activates the channel output with maximum switching frequency in excess of 50MHz. The output of this can further be tuned to compensate for amplitude offsets using a PMOS variable resistor as shown in Fig. 6. In Fig. 7, we have shown the output spectrum of Channel-0 (divisor equal to 32) of the transmitter when the 31.25MHz generated tone is switched on and off using the channel control signal with duration of 20 μ s. The reference frequency in this case is 1GHz.

Next, in order to combine the outputs of all channels and create the excitation spectrum, we use a resistor-network passive combiner followed by an output amplifier which drives the excitation coils. The maximum peak-to-peak voltage of the excitation signal for the combined channels at the combiner output with $V_{CC}=1.3V$ is 3mV, which is sufficient for driving the RF amplifiers of the MR systems. The output of the amplifier when all six channels are operating without amplitude modulation (and switching) is shown in Fig. 8 with again the reference frequency of 1GHz. The transmitter operation frequency range is from DC to 37MHz corresponding to an input reference frequency of DC to 1.2GHz, respectively. The measured offset of the channels was approximately 2.3dB which can




Fig.5: Spectrum profile of the output of the transmitter




Fig.6: Signal generation circuitry for a single channel of the transmitter.




Fig. 8: Typical output spectrum of the transmitter with reference frequency of 1GHz.




Fig.7: Output spectrum and control signal of Channel-0 in the transmitter.

be calibrated using the PMOS tuning resistor. The overall power consumption with $V_{CC}=1.3V$ varies between 6.5mW at DC to 18.2mW at 1.2GHz. The transmitter measured performances are listed in Table I.




Fig.9: The receiver circuitry for a single channel.

In the receiver, the signal induced into the receiving coils $E(t)$, which is the magnetic dipoles response to the excitation spectrum and B_0 , is initially amplified using a wideband LNA (see Fig. 9). To realize the LNA in our system, we have taken advantage of the bipolar NPN transistors of the SiGe BiCMOS process. The justification for this is the low noise characteristics, low $1/f$ noise corner, and the high gain of the NPN SiGe transistors. The LNA has a differential two-stage topology with resistive loads. The differential input of the LNA is biased internally, and there are two DC blocking external capacitors which connect the receiving coil to the input of the LNA. It is imperative to understand that the input impedance of the LNA in our design is not matched to the impedance of the coil (mostly inductive with less than 1Ω resistance) since the LNA is operating as a high input impedance amplifier not a matched amplifier typically used in RF transceivers. The LNA gain and the input referred input noise voltage are 27-31dB and $150\text{pV}/\sqrt{\text{Hz}}$ respectively at the frequency range of 10MHz to 250MHz. The LNA has an input 1dB compression point of -21dBV with overall current consumption of 19.2mA with $V_{CC}=2.9\text{V}$.

The induced signal into the coils after amplification by the LNA which amplifies the whole one octave band is introduced to the channel mixers for down conversion as shown in Fig. 3. The mixers in each channel have a CMOS active double-balanced topology with resistive loads. The required LO frequency for the mixers comes from the transmitter block. In each receiving path of all six channels there are two mixers which are driven by I and Q of that corresponding channel. Using this method we are able to achieve 23dB differential voltage gain for the mixers, with an input 1dB compression point of -15dBV. The bandwidth of the mixer which basically sets the bandwidth of the receiver chain is approximately 250kHz with the external capacitive loads.

The overall gain of the receiver is 50dB with 1dB input compression point of -42dBV. The total power consumed in the receiver is between 78.2mW and 90mW depending on the operation frequency from DC to 1.2GHz. In Table I, we have listed all the measured specifications of the system including the transmitter and receiver blocks

In our system, for narrow-band MR spectroscopy we have also included two auxiliary channels which can function independently from the MRI transceiver. The major difference between these two channels and the wide-band MRI transceiver is only in their transmitter. They do not have the initial digital dividers and their output is not combined with any other channel. Using these channels narrow-band signal excitation and coherent detection can be carried out up to 250MHz using the same LNA and mixers. The operation frequency in this mode is limited by the LNA bandwidth of the LNA and matching considerations.

IV. CONCLUSION

An integrated circuit implementation of a Spectral-Scanning MRI is shown for the first time. A coherent multiple frequency transmitter and receiver design provides high sensitivity (equivalent to a lock-in operation). We believe that this SSMRI implementation is a major step toward successful demonstration of a hand-held MRI unit.

TABLE I
SPECTRAL SCANNING MRI SYSTEM SPECIFICATIONS

SoC Specifications	
Process	IBM 8HP 0.12 μm SiGe BiCMOS
Die specifications	3mm x 3mm, 88 pins (56 I/O, 32 test)
Off-chip	MRI Coils, DC blocking capacitors
Power consumption	78.2-90mW
^1H detection	DC-37MHz, $B_0=0-0.85\text{T}$
Transmitter	
Operation frequency	MRI [DC-37MHz], NMR [DC-250MHz]
Number of channels	MRI [0:5], auxiliary [0:1]
Output level	0-3mV/channel, $I_{\text{MAX}}=2.2\text{mA}$
Modulation frequency	50MHz
In-band intermods level	-40dB of channel output
Power consumption	6.5-18.2mW for DC-1.2GHz input frequency
Receiver	
Operation frequency	<250MHz, BW=250kHz
LNA	Gain=27dB, input noise=0.15nV/Hz $^{1/2}$
Mixer	Gain=23dB, 1dB comp=-15dBV
Power Consumption	LNA=55.5mW, Mixers=16mW

REFERENCES

- [1] E. M. Haacke, R. W. Brown, M. R. Thompson, and R. Venkatesan, *Magnetic Resonance Imaging: Physical Principles and Sequence Design*, New York, Wiley, 1999.
- [2] P. Callaghan, *Principles of Nuclear Magnetic Resonance Microscopy*, New York, Oxford, 1993.
- [3] D.L. Olson *et al.*, "High-Resolution Micro-coil ^1H -NMR for Mass-Limited, Nanoliter-Volume Samples," *Science*, vol. 270, pp. 1967-1970, December 1995.
- [4] L. Ciobanu *et al.*, "3D Micron-Scale MRI of Biological Cells," *Solis Atate Nucl. Magn. Reson.*, vol 25, pp. 138-141, January 2004.
- [5] C. Massin *et al.*, "High-Q factor RF Planar Microcoils for Micro-Scale NMR Spectroscopy," *Sensors & Actuators A: Physical*, vol. 97, pp. 280-288, April 2002.
- [6] J. Perlo *et al.*, "High resolution NMR spectroscopy with a portable single-sided sensor," *Science*, pp. 1279, May 2005.
- [7] P. Prado, F. Casanova and B. Blümich, "3D imaging with a single-sided sensor: an open tomograph," *Journal of Magnetic Resonance*, vol. 166, pp.228-235, February 2004.
- [8] V.M. Kenkre, E. Fukushima and D. Sheltraw, "Simple solutions of the Torrey-Bloch equations in the NMR study of molecular diffusion," *Journal of Magnetic Resonance*, vol. 128, pp. 62-69, September 1997.
- [9] Hajimiri *et al.*, U.S. Patent App. No. 11/499,92.